

## Kinetic Spectra of the Planar Multipole Resonance Probe

Friedrichs, Michael; Oberrath, Jens

Published in: Bulletin of the American Physical Society

Publication date: 2018

**Document Version** Publisher's PDF, also known as Version of record

Link to publication

Citation for pulished version (APA): Friedrichs, M., & Oberrath, J. (2018). Kinetic Spectra of the Planar Multipole Resonance Probe. *Bulletin of the American Physical Society*, 2018, Article TF3.00007. http://meetings.aps.org/Meeting/GEC18/Session/TF3.7

## **General rights**

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
  You may freely distribute the URL identifying the publication in the public portal ?

## Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Abstract Submitted for the GEC18 Meeting of The American Physical Society

Kinetic Spectra of the Planar Multipole Resonance Probe<sup>1</sup> MICHAEL FRIEDRICHS, Institute of Product and Process Innovation, Leuphana University, JENS OBERRATH, Institute of Product and Process Innovation, Leuphana University Lneburg, Germany — The planar multipole resonance probe is suitable for industrial plasma diagnostic purposes and consists of two half-disc electrodes, which can be integrated into the chamber wall of the reactor. Due to its minimal invasive character, inner plasma parameters can be monitored during the process. Based on a fluid model it is possible to determine the electron density from a detected resonance frequency in the measured spectrum. However, to monitor also the electron temperature an additional resonance parameter, e.g. the half-width of the resonance peak, is necessary. It is strongly influenced by kinetic effects, therefore a study of a kinetic model to obtain a relation between the half-width and the electron temperature is required. In this work such a kinetic model based on functional analytic methods and first spectra are presented.

<sup>1</sup>The authors gratefully acknowledge funding by the German Research Foundation (DFG) within the project OB 469/1-1

Michael Friedrichs Institute of Product and Process Innovation, Leuphana University

Date submitted: 14 Jun 2018

Electronic form version 1.4